



राष्ट्रीय पादप जीनोम अनुसंधान संस्थान

(जैव प्रौद्योगिकी विभाग, विज्ञान एवं प्रौद्योगिकी मंत्रालय, भारत सरकार का स्वायत्त अनुसंधान संस्थान)

NATIONAL INSTITUTE OF PLANT GENOME RESEARCH

(An Autonomous Institution of the Department of Biotechnology, Ministry of Science and Technology, Government of India)

अरुणा आसफ अली मार्ग, पो. बाक्स नं. 10531, नई दिल्ली-110067

Aruna Asaf Ali Marg, Post Box Number 10531, New Delhi-110067

No. 8/7/2018-19/NIPGR/S&P

Dated 08/10/2018

Corrigendum

This has reference to our tender no. 8/7/2018-19/NIPGR/S&P dated 5/9/2018 published vide CPPP Tender ID No. 2018_NIPGR_379795_1 regarding supply and installation of 04 Nos. of Plant Growth Chambers at our Institute.

In this context, the following changes have been made in the specifications of above mentioned Chambers.

1. Point no 2: Light Intensity: 1100 micromoles/sqm/sec (below 6" of the lamp) in single tier Configuration. Light Intensity: 500 micromoles/sqm/sec on each of the two tiers in 2-Tier configuration. Both having horizontally placed lamps.
2. Point no. 4: Temperature: +4 Deg C to 40 Deg C with Lights OFF and +10 Deg C to 40 Deg C with Lights ON.
3. Point No. 9: Dimensions: Approx. 100 inches (L) x 36 inches (W) x 101 inches (H).
4. Point No. 14: Alerts: Email Alert only.
5. Include Air Flow in the specs as it is a very important parameter for Plant Growth. Uniformly distributed upward (in single tier) and horizontal (in two tier) Air Flow.
6. Point no. 19: Cabinet Voltage Please include 400V 3-phase 50Hz for Cabinet operation.
7. Point no. 20: Performance and installation: Performance statement of 5 such installations in India out of which at least 01 should be working satisfactorily for more than 5 years.
8. Point No. 21: Stabilizer and UPS: Suitable servo voltage stabilizer for chamber and offline UPS for controller only.

Last Date/ time of submission of Bids: 16/10/2018 (3.00 P.M)

Vice

11/10/2018 (3.00 P.M)

Opening of Bids: 17/10/2018 (3.00 P.M)

Vice

12/10/2018 (3.00 P.M)

This Corrigendum will form part of the Original Tender Document. All other terms and conditions of the above mentioned tender remain unchanged.

(क्रय एवं भंडार अधिकारी)